Via della Meccanica, 25, 41012 Carpi (MO) - Italy
+39 059 653274 - info@giorgiobormac.com

XS Basic pH 9.00 /25°C (blue), 250 ml autocal bottle Test solution



Code: *51100053*

TECHNICAL DATA SHEET

Features

Accuracy	± 0.02
Certificate	Yes
Colour	Blue
Composition	Boric Acid, Potassium Chloride and Sodium Hydrate
Package	250 ml with integrated calibration compartment
Description	±0.02 / 25°C - Blue
Name	pH 9.00
Nominal value	pH 9.00 @ 25 °C
Volume	250 ml
hazards	Non-hazardous. No ADR restrictions