Via della Meccanica, 25, 41012 Carpi (MO) - Italy
+39 059 653274 - info@giorgiobormac.com

XS Basic pH 4.01 /25°C (red), 250 ml autocal bottle Test solution



Code: *51100033*

TECHNICAL DATA SHEET

Features

Accuracy	± 0.02
Certificate	Yes
Colour	Pink
Composition	Potassium phthalate acid, potassium chloride
Package	250 ml with integrated calibration compartment
Description	±0.02 / 25°C - Red
Name	pH 4.01
Nominal value	pH 4.01 @ 25 °C
Volume	250 ml
hazards	Non-hazardous. No ADR restrictions