Via della Meccanica, 25, 41012 Carpi (MO) - Italy
+39 059 653274 - info@giorgiobormac.com

XS Professional pH 9.21 /25°C, 500 ml autocal bottle Calibration solution



Code: *51300123*

TECHNICAL DATA SHEET

Features

Accuracy	± 0.02
Certificate	Yes
Colour	Blue
Composition	Sodium borate (CAS 1303-96-4), blue dye
Package	500 ml with integrated calibration compartment
Description	±0.02 / 25°C - Blue
Name	pH 9.21
Nominal value	pH 9.21 @ 25 °C
Volume	500 ml
hazards	Non-hazardous. No ADR restrictions