Via della Meccanica, 25, 41012 Carpi (MO) - Italy
+39 059 653274 - info@giorgiobormac.com

XS Basic pH 10.01 /25°C (colourless), 500 ml autocal bottle Test solution



Code: *51100163*

TECHNICAL DATA SHEET

Features

Accuracy	± 0.02
Certificate	Yes
Colour	Colourless
Composition	Sodium carbonate, sodium bicarbonate
Package	500 ml with integrated calibration compartment
Description	±0.02 / 25°C - Colourless
Name	pH 10.01
Nominal value	pH 10.01 @ 25 °C
Volume	500 ml
hazards	Non-hazardous. No ADR restrictions